



**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2016

Biomechanical evaluation of adjunctive cerclage wire fixation for the prevention of periprosthetic femur fractures using cementless press-fit total hip replacement

Christopher, Scott A ; Kim, Stanley E ; Roe, Simon ; Pozzi, Antonio

DOI: <https://doi.org/10.1016/j.tvjl.2016.04.014>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-127148>

Journal Article

Accepted Version



The following work is licensed under a Creative Commons: Attribution-NonCommercial-NoDerivatives 4.0 International (CC BY-NC-ND 4.0) License.

Originally published at:

Christopher, Scott A; Kim, Stanley E; Roe, Simon; Pozzi, Antonio (2016). Biomechanical evaluation of adjunctive cerclage wire fixation for the prevention of periprosthetic femur fractures using cementless press-fit total hip replacement. *Veterinary Journal*, 214:7-9.

DOI: <https://doi.org/10.1016/j.tvjl.2016.04.014>

1 **Short Communication**

2

3

4 **Biomechanical evaluation of adjunctive cerclage wire fixation for prevention of**
5 **periprosthetic femur fractures using cementless press-fit total hip replacement.**

6

7

8 Scott A. Christopher^a, Stanley E. Kim^a, Simon Roe^b, Antonio Pozzi^{a*}

9

10

11 ^a Comparative Orthopaedics Biomechanics Laboratory, College of Veterinary Medicine,
12 University of Florida, Gainesville, FL

13 ^b Randall B. Terry Companion Animal Veterinary Medical Center, College of Veterinary
14 Medicine, North Carolina State University

15 * Corresponding author

16

17

18 Email Addresses:

19 Scott A. Christopher: sachristopher@ufl.edu

20 Stanley E. Kim: stankim@ufl.edu

21 Simon Roe: simon_roe@ncsu.edu

22 Antonio Pozzi: antonio.pozzi@uzh.ch

23

24 **Abstract**

25 Periprosthetic femoral fractures are a common complication associated with cementless press-fit
26 total hip arthroplasty. The use of prophylactic cerclage wire fixation has been advocated to
27 reduce this complication. The objective of this study was to evaluate whether a double loop
28 cerclage wire, used as adjunctive fixation, increases the peak torsional load to failure in femora
29 implanted with press-fit cementless stems. Peak torsional load to failure was compared between
30 femora without adjunctive fixation and femora receiving a 1 mm double loop cerclage wire
31 placed proximally to the lesser trochanter. Femora treated with adjunctive cerclage wire fixation
32 failed at 20% greater peak torque ($P = 0.0001$). In conclusion, a double loop cerclage wire may
33 aid in the prevention of periprosthetic fractures associated with press-fit cementless femoral
34 stems.

35

36 Keywords: total hip arthroplasty, cerclage, press fit, periprosthetic fracture

37

38 The BioMedtrix Biologic Fixation (BFX) total hip replacement (THR) system is a
39 commonly used system for cementless press-fit THR in dogs (Marcellin-Little et al., 1999).
40 Excellent outcomes have been reported for this system, with high client satisfaction and full
41 return to function in most cases (Lascelles et al., 2010). One complication of BFX THR includes
42 periprosthetic femoral fracture (Ganz et al., 2010) with a reported prevalence of 5% (Hunt and
43 Preston, 2009). Postoperative femoral fractures are typically spiral, extending from the base of
44 the femoral neck distally, suggesting that the most likely mechanism may be a combination of
45 axial and torsional loads (Ganz et al., 2010; Jasty et al., 1992; Jasty et al., 1994). Prophylactic
46 cerclage wire has been advocated for preventing periprosthetic fractures in dogs (Ganz et al.,
47 2010). The goal of this study was to biomechanically evaluate the prophylactic use of double
48 loop cerclage wire as adjunctive fixation for BFX THR.

49
50 This study was approved by the institutional animal care and use committee (#20138104,
51 September 2013). Paired femora without evidence of pathology were collected from 10 mature
52 dogs, euthanized for unrelated reasons. After taking orthogonal radiographs of the femur, the
53 canal flare index, defined as the ratio between width of the medullary canal at the lesser
54 trochanter and at the narrowest point of the isthmus (the region where the canal narrows distal to
55 the proximal femoral metaphysis), was measured using the craniocaudal radiographic views
56 (DeYoung et al., 1993). Femora with a canal flare index < 1.8 (stovepipe morphology) were
57 excluded (Ganz et al., 2010). Only femora that fitted a size 8 BFX stems (Biomedtrix, LLC)
58 were selected.

59

60 Femora were prepared and stems were inserted as previously described (McCulloch et al.,
61 2012). Paired femora were randomly assigned to either the treatment group (with adjunctive
62 cerclage wire) or the control group (without cerclage wire). For the treatment group, one double-
63 loop cerclage wire of 1 mm (selected based on previous recommendations (McCulloch et al.,
64 2010) was applied proximal to the lesser trochanter. Mean **cranial** canal fill was calculated on
65 craniocaudal images by averaging the canal fill at the smooth/porous junction (proximal third) of
66 the BFX stem, 5 mm proximal to the distal tip of the implant, and midway between these two
67 points; **mean lateral** canal fill was measured in the lateral projections 5 mm above the **distal tip of**
68 **the implant** in the lateral projection (DeYoung et al., 1993).

69
70 The distal end of the femur was potted (3M Bondo, 3M Corp) using a custom jig and
71 fixed to the base of a servohydraulic materials testing machine (MTS 858 Mini Bionix II)
72 equipped with an axial-torsional load cell. Specimens were mounted in the MTS and a custom
73 made jig was attached to the actuator of the MTS (Pozzi et al., 2013). After biomechanical
74 preconditioning, specimens were loaded to 300 N of axial compression and 0.5 Nm of torsional
75 load for 2 seconds, then loaded to failure under a constant axial load of 300 N and internal
76 rotation of the femur applied at a rate of 10°/s. Load at failure was defined as the load at which
77 the fracture was first initiated and was associated with a sudden decrease in load and torque.
78 High definition digital video camera was used to record fracture propagation. Fractures were
79 classified using the Vancouver classification scheme (Brady et al., 2000).

80
81 For data analysis, a Shapiro-Wilk test was used to evaluate the normality of distribution.
82 Canal flare index, canal fill, and peak torque load to failure were compared between the

83 treatment and control group using paired t-tests (Prism, GraphPad Software, Inc.) $P < 0.05$ was
84 considered significant. Slow motion and frame-by-frame video analysis was used to characterize
85 the location and initiation site of the periprosthetic fracture.

86 Canal flare index and mean **cranial and lateral** canal fill was not significantly different
87 between the treatment and control groups. Cerclage wire application resulted in a significant
88 increase in peak torque **at failure** of $19.6\% \pm 1.2\%$.

89
90 All control femora failed with a similar fracture configuration, consistent with a
91 Vancouver class B₂ fracture (Fig 1A). Femora with cerclage wire fixation fractured by one of
92 three distinct manners: One femur fractured through the region at the base of the femoral neck
93 resulting in cerclage wire failure and a loose femoral stem (class B₂); 2 femora failed with a
94 spiral fracture extending below the stem into the diaphysis, (class C); the remaining 7 femora
95 resulted in fractures originating as a small non-displaced fissure in region at the base of the
96 femoral neck propagating distally resulting in a fracture with a well-fixed femoral stem (class B₁)
97 (Fig. 1B).

98
99 This study showed that a double loop cerclage wire placed proximally to the lesser
100 trochanter significantly increased the resistance to periprosthetic fractures in femora implanted
101 with BFX under combined axial and torsional loads. Based on these results, prophylactic
102 cerclage wire may reduce the likelihood of periprosthetic femoral fractures associated with BFX
103 *in vivo*. This is in accordance with results of both clinical and mechanical studies investigating
104 the benefit of cerclage wire application in conjunction with BFX (Ganz et al., 2010; McCulloch
105 et al., 2012).

106 The fracture configurations observed in this study were consistent with previous reports
107 of periprosthetic fractures (Ganz et al., 2010). As expected, a single cerclage wire limited fissure
108 widening proximally, but had no protective effect distally. Therefore, multiple cerclage wires
109 may be a better choice in dogs at high risk of fractures because of the ability to control fracture
110 propagation distal to the lesser trochanter [as suggested by previous human studies \(Fishkin et al.](#)
111 [1999\)](#).

112
113 One limitation of this study may have been the variability associated with femoral
114 preparation and implant application. However, this variation is considered insignificant because
115 digital templating of stem size was performed by two experienced surgeons (SK and AP),
116 surgical procedures were performed by a single surgeon (SK), and mean [cranial and lateral](#) canal
117 fill was not significantly different between treatment and control groups.

118
119 In conclusion, application of prophylactic cerclage wire should be considered at the time
120 of stem implantation in dogs with a proximal femoral morphology suitable for BFX but at risk of
121 periprosthetic fractures.

122

123

124

125 **Conflict of Interest Statement**

126 None of the authors of this paper has a financial relationship with other people or organizations
127 that could inappropriately influence or bias the content of the paper. Antonio Pozzi is an
128 instructor for THR Biomedtrix courses.

129

130 **Acknowledgements**

131 This work was presented at the Fourth World Veterinary Orthopedic Congress on March 4, 2014
132 in Breckenridge, Colorado. The femoral stem prostheses used in the study were donated by
133 Biomedtrix, LLC.

134

135 **References**

136 Brady, O.H., Garbuz, D.S., Masri, B.A., Duncan, C.P., 2000. The reliability and validity of the
137 Vancouver classification of femoral fractures after hip replacement. *The Journal of*
138 *Arthroplasty* 15, 59–62.

139

140 DeYoung, D.J., Schiller, R.A., DeYoung, B.A., 1993. Radiographic assessment of a canine
141 uncemented porous-coated anatomic total hip prosthesis. *Veterinary Surgery* 22, 473–
142 481.

143

144 [Fishkin Z., Han S.M., Ziv I., 1999. Cerclage wiring technique after proximal femoral fracture in
145 total hip arthroplasty. *Journal of Arthroplasty* 14, 98-101.](#)

146

147 Ganz, S.M., Jackson, J., VanEnkevort, B., 2010. Risk factors for femoral fractures after canine
148 press-fit cementless total hip arthroplasty. *Veterinary Surgery* 39, 688–695.

149

150 Hunt, L., Preston, C., 2009. Retrospective correlation study between femoral geometry and
151 femoral component complications after Biomedtrix BFX cementless THR. In:
152 *Proceedings of The Australian & New Zealand College of Veterinary Scientists Science*
153 *Week, Queensland, Australia, pp.9-10.*

154

155 Jasty, M., Bragdon, C.R., Rubash, H., Schutzer, S.F., Haire, T., Harris, W., 1992. Unrecognized
156 femoral fractures during cementless total hip arthroplasty in the dog and their effect on
157 bone ingrowth. *The Journal of Arthroplasty* 7, 501–508.

158

- 159 Jasty, M., O'Connor, D.O., Henshaw, R.M., Harrigan, T.P., Harris, W.H., 1994. Fit of the
160 uncemented femoral component and the use of cement influence the strain transfer the
161 femoral cortex. *Journal of Orthopaedic Research* 12, 648–656.
162
- 163 Lascelles B.D., Freire M., Roe S.C., DePuy V., Smith E., Marcellin-Little D.J., 2010. Evaluation
164 of functional outcome after BFX total hip replacement using a pressure sensitive
165 walkway. *Veterinary Surgery* 39, 71-77.
166
- 167 Marcellin-Little, D.J., DeYoung, B.A., Doyens, D.H., DeYoung, D.J., 1999. Canine uncemented
168 porous-coated anatomic total hip arthroplasty: results of a long-term prospective
169 evaluation of 50 consecutive cases. *Veterinary Surgery* 28, 10–20.
170
- 171 McCulloch, R.S., Roe, S.C., Marcellin-Little, D.J., Mente, P.L., 2012. Resistance to subsidence
172 of an uncemented femoral stem after cerclage wiring of a fissure. *Veterinary Surgery* 41,
173 163–167.
174
- 175 Pozzi, A., Peck, J.N., Chao, P., Choate, C.J., Barousse, D., Conrad, B., 2013. Mechanical
176 evaluation of adjunctive fixation for prevention of periprosthetic femur fracture with the
177 zurich cementless total hip prosthesis. *Veterinary Surgery* 42, 529–534.
178

179 **Table 1**

180 Mean \pm SD for the treatment (with adjunctive cerclage wire) and control groups and
181 corresponding P-values (treatment vs. control) for the parameters canal flare index, mean cranial
182 canal fill (measured on craniocaudal radiographs), lateral canal fill (measured on lateral
183 radiographs) and peak torque at failure.

	Treatment	Control	P-value
Canal flare index	2.09 \pm 0.167	2.10 \pm 0.155	0.39
Mean cranial canal fill	75.2 \pm 6.67%	75.7 \pm 5.8%	0.44
Mean lateral canal fill	65.4 \pm 6.28%,	65.76 \pm 6.83%	0.69
Peak torque at failure	22.71 \pm 2.41 Nm	18.98 \pm 2.91Nm	0.0001

184

185

186

187 **Figure legends**

188 Figure 1: Paired specimen representative of the typical fracture patterns noted in implanted
189 femora with and without cerclage wire. In the untreated femur (A) the stem subsided (black
190 arrows) and caused a Vancouver B₂ fracture: a fissure line originated from the base of the
191 femoral neck and propagated distally resulting in an unstable femoral stem. In the contralateral
192 femur with a double cerclage wire applied proximally to the lesser trochanter, a small non-
193 displaced fissure at the base of the femoral neck with a well-fixed femoral stem can be noted
194 (class B₁).