**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2014

Improved localization of implanted subdural electrode contacts on magnetic resonance imaging with an elastic image fusion algorithm in an invasive electroencephalography recording

Stieglitz, L H ; Ayer, C ; Schindler, K ; Oertel, M F ; Wiest, R ; Pollo, Claudio

Abstract: BACKGROUND: Accurate projection of implanted subdural electrode contacts in presurgical evaluation of pharmacoresistant epilepsy cases by invasive electroencephalography is highly relevant. Linear fusion of computed tomography and magnetic resonance images may display the contacts in the wrong position as a result of brain shift effects. OBJECTIVE: A retrospective study in 5 patients with pharmacoresistant epilepsy was performed to evaluate whether an elastic image fusion algorithm can provide a more accurate projection of the electrode contacts on the preimplantation magnetic resonance images compared with linear fusion. METHODS: An automated elastic image fusion algorithm (AEF), a guided elastic image fusion algorithm (GEF), and a standard linear fusion algorithm were used on preoperative magnetic resonance images and postimplantation computed tomography scans. Vertical correction of virtual contact positions, total virtual contact shift, corrections of midline shift, and brain shifts caused by pneumocephalus were measured. RESULTS: Both AEF and GEF worked well with all 5 cases. An average midline shift of 1.7 mm (SD, 1.25 mm) was corrected to 0.4 mm (SD, 0.8 mm) after AEF and to 0.0 mm (SD, 0 mm) after GEF. Median virtual distances between contacts and cortical surface were corrected by a significant amount, from 2.3 mm after linear fusion algorithm to 0.0 mm after AEF and GEF ($P < .001$). Mean total relative corrections of 3.1 mm (SD, 1.85 mm) after AEF and 3.0 mm (SD, 1.77 mm) after GEF were achieved. The tested version of GEF did not achieve a satisfying virtual correction of pneumocephalus. CONCLUSION: The technique provided a clear improvement in fusion of preimplantation and postimplantation scans, although the accuracy is difficult to evaluate.

DOI: <https://doi.org/10.1227/NEU.0000000000000473>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-107687>

Journal Article

Accepted Version

Originally published at:

Stieglitz, L H; Ayer, C; Schindler, K; Oertel, M F; Wiest, R; Pollo, Claudio (2014). Improved localization of implanted subdural electrode contacts on magnetic resonance imaging with an elastic image fusion algorithm in an invasive electroencephalography recording. *Neurosurgery*, 10(4):506-512.

DOI: <https://doi.org/10.1227/NEU.0000000000000473>

**Improved localization of implanted subdural electrode contacts on
MRI using an elastic image fusion algorithm in invasive EEG
recording**

Stieglitz, Lennart Henning ¹

Ayer, Christian²

Schindler, Kaspar³

Oertel, Markus Florian⁴

Wiest, Roland⁵

Pollo, Claudio⁴

¹Department of Neurosurgery, Zürich University Hospital, University of Zürich, Zürich, Switzerland

²University of Bern, Bern, Switzerland

³Department of Neurology, Inselspital, Bern University Hospital and University of Bern, Bern, Switzerland

⁴Department of Neurosurgery, Inselspital, Bern University Hospital and University of Bern, Bern, Switzerland

⁵Diagnostic and Interventional Neuroradiology, Inselspital, Bern University Hospital and University of Bern, Bern, Switzerland

ABSTRACT

Background:

Accurate projection of implanted subdural electrode contacts in presurgical evaluation of pharmacoresistant epilepsy cases by invasive EEG is highly relevant. Linear fusion of CT and MRI images may display the contacts in the wrong position due to brain shift effects.

Objective:

A retrospective study in five patients with pharmacoresistant epilepsy was performed to evaluate whether an elastic image fusion algorithm can provide a more accurate projection of the electrode contacts on the pre-implantation MRI as compared to linear fusion.

Methods:

An automated elastic image fusion algorithm (AEF), a guided elastic image fusion algorithm (GEF), and a standard linear fusion algorithm (LF) were used on preoperative MRI and post-implantation CT scans. Vertical correction of virtual contact positions, total virtual contact shift, corrections of midline shift and brain shifts due to pneumocephalus were measured.

Results:

Both AEF and GEF worked well with all 5 cases. An average midline shift of 1.7mm (SD 1.25) was corrected to 0.4mm (SD 0.8) after AEF and to 0.0mm (SD 0) after GEF. Median virtual distances between contacts and cortical surface were corrected by a significant amount, from 2.3mm after LF to 0.0mm after AEF and GEF ($p<.001$). Mean total relative corrections of 3.1 mm (SD 1.85) after AEF and 3.0mm (SD 1.77) after GEF were achieved. The tested version of GEF did not achieve a satisfying virtual correction of pneumocephalus.

Conclusion:

The technique provided a clear improvement in fusion of pre- and post-implantation scans, although the accuracy is difficult to evaluate.

Running Title: Elastic fusion in invasive recording

Keywords: EEG, elastic image fusion, epilepsy, invasive recording

INTRODUCTION

In pharmacoresistant epilepsy cases, the decision to perform invasive investigation using intracranial electroencephalography (EEG) electrodes in order to localize the epileptogenic locus is often one of the last diagnostic options. In such cases, patients and physicians are driven by the hope of identifying a region of the brain whose resection might lead to significantly improved seizure control or even complete seizure freedom and increased quality of life. There are currently several accepted surgical methods to place electrode contacts on the brain surface to identify such regions.

The first method involves performing an extensive craniotomy and positioning a multi-contact grid on the cortex.¹ The second but less invasive method involves stereotactic implantation of multiple electrodes inside the brain through a series of trepanations, called stereo-EEG.^{2,3} The third and least invasive method is the implantation of 4 to 10 contact strip-electrodes on the cortex through a small number (at least one per side) of limited trepanations.⁴ After electrode implantation, the direct positioning of the contacts on the cortex allows EEG recording with excellent signal-to-noise ratios and minimal muscle artifacts over several days, and leads to a much better delineation of the epileptogenic zone as compared to extracranially recorded EEG.⁵ These EEG data can then be used for planning a potentially curative resective surgery, with the aim to achieve freedom from seizures.

All techniques enable exact localization of the contact position on the cortex, which is crucial for interpretation of the EEG signals. However, an imaging technology that allows visualization of metal electrode contacts and brain tissue simultaneously would be preferable. Unfortunately MRI can cause tissue damage due to energy transfer into heat.⁶ Furthermore, the contacts cause artifacts in most imaging sequences that are far more intensive than the signal of the contacts themselves. On the other hand, in a CT scan the contacts can be visualized easily and even without image distortion, but due to artifacts around the contacts and inferior quality of soft-tissue contrast, localization of the contacts in relation to the brain surface is also difficult.⁷ Combination of both techniques using image fusion of preoperatively acquired MRI with postoperative CT scans is common and provides the required spatial information, but is susceptible to the effects of brain shift. Cerebrospinal fluid (CSF) loss during electrode

implantation and air trapped subdurally may cause a considerable dislocation of cortical structures.^{8,9} A linear image fusion, allowing only translations, rotations, scaling and skewness to align two image datasets, might localize the contacts in the wrong position.^{10,11} Elastic image fusion algorithms are a relatively new development and are not yet standard in commercial software used in neurosurgery. In addition to linear translations they allow local modifications of the image datasets to achieve a better alignment. They may be helpful for solving this problem as they theoretically can compensate for these brain shift effects and enable visualization of the actual contact positions on the gyri. To analyze the preliminary results of such an innovative approach, a retrospective study in five patients with pharmacoresistant epilepsy was performed to evaluate whether an elastic image fusion algorithm can provide a more accurate projection of the electrode contacts on the pre-implantation MRI than simple elastic fusion.

METHODS

Patients

Twenty-three patients (6 males, 17 females) suffering from pharmacoresistant epilepsy underwent an invasive recording phase between January and December 2011 at the Bern University Hospital (Inselspital). All patients had previously undergone a non-invasive recording phase, but the epileptogenic brain areas could not be as precisely localized as required to directly proceed to resective surgery.

Application of the elastic image fusion required a complete imaging dataset, which is described in detail below. A subgroup of five patients, two males and three females with an average age of 19 years (SD 2.9), ultimately fulfilled these requirements.

Pre- and postoperative neuroimaging

For the stereotactic implantation of additional hippocampal depth electrodes¹² all patients received pre- and postoperative CT scans. Preoperatively, the patients underwent a T2 MRI scan (T2 weighted spin echo sequence: TR=2200 ms, S=1 mm (gap 0), FOV=256 mm, matrix=256×256 on a 3 T Magnetom Verio MR system, Siemens Healthcare, Erlangen, Germany) and a native CT scan (tube current = 180 mA, kvp = 120 kV, standard kernel, slice thickness 1 mm (supratentorial/infratentorial), and FOV = 220 mm on a GE Lightspeed 8-row detector scanner,

GE Healthcare, Milwaukee, Wisconsin, USA). Postoperatively (day of surgery) all patients received an additional native CT scan as described above.

Electrode implantation technique

To reduce morbidity and to allow maximal coverage of the brain surface we implanted multiple 4 to 6-contact strip electrodes in a star-like manner from frontal and frontotemporal via 14-mm-trepanations. To minimize the risk of infection the patient's head was shaved and iv antibiotics were given perioperatively. The ideal positions of the trepanations were localized using a neuronavigation system (VectorVision2, BrainLab, Feldkirchen, Germany). Under general anesthesia, trepanation was performed and the dura was opened. Four to eight contact strip electrodes (Ad-Tech medical instrument corporation, Racine, USA) were implanted under neuronavigation and fluoroscopic guidance (Table 1). Depending on the clinical presentation of the patient one burr hole was placed uni- or bilaterally over the Sylvian fissure with 3 to 4 strips going around the temporal lobe and 2 to 4 strips covering the frontal and frontoparietal lobes. To reduce CSF loss, the trepanation was sealed using fibrin glue after implantation of the final electrode. The cables were subcutaneously tunneled and externalized through the skin at least 4 cm distant from the trepanation site to reduce infection. Postoperatively, the position of the electrode contacts and absence of potential subdural hematomas were confirmed on a native CT scan.

Image co-registration

The most common method used for co-registration of two image datasets is the so-called mutual information (MI) method. It originates from the information theory and measures the statistical dependency between two datasets and was shown to perform well in co-registration of CT and MRI.^{13,14} In linear co-registration the alignment of the two datasets is achieved by translation, rotation, scaling and skewness, each along three degrees of freedom (DOF) adding up to 12 DOF altogether. Elastic co-registration should be able, in addition to the linear translation, to apply changes only locally and thus deliver more precise results, especially when the image acquisition time is different or even when pre- and postoperative datasets are to be fused.

Linear image fusion

All image fusions were performed using a pre-release version provided by Brainlab iPlan (Brainlab, Feldkirchen, Germany). The algorithm for linear image fusion (LF) applies 12 DOF and is identical with that included in the commercial software version iPlan 3.0. The region of interest is automatically defined and covers the entire skull of the patient. In all fusion procedures the postoperative CT scan was fused onto the preoperative MRI scan.

Elastic image fusion

The elastic image fusion was performed using two different versions of the elastic fusion algorithm. The first one, referred to as automated elastic fusion (AEF), tries to find a corresponding position in the first dataset for each structure in the second. The algorithm runs fully automated.

The second one, referred to as guided elastic fusion (GEF), allows manual definition of corresponding structures in both images; during the automated fusion process, these manually-defined structures may have no counterpart in the other dataset. Examples for such structures would be tumors in pre- and postoperative images or intracranial air, which are not present in preoperative images. In the present study we segmented intracranial air for GEF.

Both algorithms allow deformation of an underlying grid in three dimensions to achieve an optimal fit of the anatomical structures. Figure 1 shows a deformation map indicating direction and intensity of deformation for each position of the grid. Just like in the linear fusion processes, the postoperative CT scans were elastically fused onto the preoperative MRI.

Assessment of the effect of elastic image fusion on virtual brain shift

CSF loss and resulting pneumocephalus often lead to a lateral shift of the midline to the contralateral side. Therefore, as the most straightforward parameter, the lateralization of the midline at the level of the foramen of Monroe was measured in millimeters both after linear and after elastic image fusion using the two different software versions, AEF and GEF.

Second, the size of frontally trapped subdural air was judged in the postoperative CT after linear and after elastic image fusion. The maximum thickness in an axial slice and the volume of trapped air were measured.

After linear image fusion the electrode contacts on CT are sometimes virtually projected inside the cortex or even subcortically. An optimal fusion should enable visualization of the electrode contacts correctly on the cortical surface and should compensate for brain shift effects. To examine the ability of both elastic fusion algorithms (AEF and GEF) to achieve this, we measured the distances of the virtual contacts' positions and the brain surface, and compared the results with the position of the contacts after linear image fusion. The distances are given in millimeters and measured separately in the frontal, parietal and temporal regions. Furthermore, the shortest distance between electrode contacts after linear and elastic image fusion were measured to judge the dimension of the elastic fusion's effect.

Statistics and ethics

All patient data were anonymized before import into the iPlan software. The local ethics committee approved this retrospective analysis. Statistical analyses included Welch two-sample t-test and Wilcoxon non-parametric test using "R" statistics programming language.¹⁷ A p-value of less than .05 was considered significant. All values were not normally distributed (Shapiro-Wilk Normality test).

RESULTS

Midline shift correction

Correction of the midline shift was one of the major parameters used to evaluate elastic fusion's accuracy. As a point of reference to measure this shift we defined an axial slice through the septum pellucidum directly above the foramen of Monro. While the median midline shift in the postoperative CT (evaluated by LF) was 1.9 mm, it was corrected to a median of 0 mm by both AEF and GEF. Due to the small number of patients included in the study the result was not statistically significant (Table 2).

Pneumencephalus correction

The second task for elastic fusion algorithms was the correction of pneumencephalus in the fused images. Neither AEF nor GEF corrected sufficiently for this problem, as shown for measurements of frontal air thickness (Table 3) and air volumetry (Table 4). Therefore, the elastic fusion algorithms used in the present study ultimately provided no significant reduction of pneumencephalus.

Electrode contact position

Both AEF and GEF achieved an effective reduction of the electrode – cortex-distance, as shown in Figure 2. This result was achieved in all tested areas: frontal, temporal and parietal regions. The reduction of distance compared to the linear fusion was statistically significant (Wilcoxon; $p < .001$; Table 5).

Relative virtual electrode contact position correction

Elastic image fusion led to relevant virtual shifts of the electrode contact positions by median distances of 2.9 mm (AEF) and 2.8 mm (GEF) respectively (Table 5). Differences between the two fusion algorithms tested in this study were not statistically significant. However, statistically significant regional differences were found for the frontal versus temporal cortices ($p = .031$) and the frontal versus parietal cortices ($p < .001$).

DISCUSSION

Invasive recording workup for patients suffering from pharmacoresistant epilepsy is often a stressful and demanding procedure for both the patient and the attending physician. Though the procedure can be considered safe, effective and a gold-standard, implantation of electrodes into the skull and brain exclusively for diagnostic purposes is an option that only patients desperate for treatment of their seizures will undergo. This makes it especially important that everything is done to make the results optimally useful for treatment planning. Accurate projection of the implanted electrode contacts on the brain surface is a crucial step, especially when it involves resection planning in or close to eloquent cortex. With grid implantation through craniotomy, the brain surface can be photographed and images can be superimposed onto a cortical MRI reconstruction.^{18–23} Combinations of photography and 2D radiography have been described,¹⁰ as have linear superpositions using a 3D visualization system.⁹ Contacts under the margins of the craniotomy or placed through burr-holes cannot be documented photographically; these contacts are also subject to brain shift, which can be corrected mathematically.^{8,23}

Using the anatomical information of the CT scan and combining this with the high spatial and contact soft-tissue resolution of MRI provides a potential new option to handle this situation. Modern computers allowing highly complex calculations, together with newly developed elastic

fusion algorithms, can be used to compensate for brain shift effects in the future. The algorithms tested and presented here performed the fusion in 5 to 10 minutes per case on a BrainLab iPlan server (HP ProLiant DL360p Gen 8: 2xIntel Xeon E5-2667, 2.9 GHz, 32GB RAM, 4x300 GB HDD). One possible use is the fusion of postoperative CT after subdural electrode implantation with preoperative MRI for contact superposition.

Display of electrode contacts on the brain surface

Linear fusion of pre- and postoperative images often localizes the position of the electrode contacts incorrectly, often inside the cortex or even subcortically (Figures 3A and 3C). Automated elastic image fusion (AEF) led to a display of the contacts that was clearly distant from where they were shown after linear fusion. The effect was stronger in the temporal than in the frontal region, and also affected contacts implanted parietally. Some of the subgroups of contacts in different brain regions are very small and allow only limited interpretation.

After elastic fusion (both AEF and GEF) the contacts were displayed clearly closer to the brain surface (Figures 3B and 3D). However, a clear superiority of the GEF over the AEF could not be shown.

Plausibility of the results

Because there are usually few landmarks visible on both CT and MRI to evaluate the quality of the elastic fusion, correction of major shift effects was used to validate the fusion results. The interhemispheric midline should normally be located in the middle of the skull. However, shifts to either side can occur after trepanation of the skull, loss of CSF, intracranial trapping of air, brain swelling, and resection of space occupying lesions or epileptogenic brain tissue. Both elastic fusion algorithms completely compensated for the midline shift.

Originally our purpose was to develop a simple automated fusion algorithm. Unfortunately, the elastic fusion algorithms were not accurate enough to compensate for the shift caused by pneumocephalus. The finding that this algorithm did not sufficiently deal with additional volumes or non-existing volumes in either of the fused images led to the idea of creating an additional guided algorithm. The GEF has the advantage that it utilizes corresponding points or structures in both image sets to support the automated structure detection. Furthermore, a function was added to define a volume in one of the image sets that might undergo major change

or might not be present in the other image set. Potential examples therefore might be a tumor removed, a hematoma evacuated, or intracranial air inoculated.

Validation of elastic image fusion results is very difficult in cases where there is no way to document the electrode contact positions visually. A photographic documentation of the electrode contact positions would have been preferable -- as shown by Tao et al,²⁰ LaViolette et al,²¹ and Pieters et al,²² in cases where a grid was implanted through a craniotomy. Dykstra et al,²³ faced the same problem as we do with localizing electrode contacts that are implanted through burr-holes. Not being able to compare with intraoperative photographs, they simply shift the electrode contacts to the level of the pia.

CONCLUSION

We demonstrated that elastic fusion can produce more plausible results than the linear fusion, but the available algorithms must be developed further to deal with complex situations, such as fusing pre- and postoperative images.

The technique is fast, fully automated, and offers high flexibility. The scope of possible applications of the presented software is not limited to epilepsy surgery. Another possible field of application would be fusion of pre- and postoperative images after tumor resection to better identify the position of tumor remnants in the preoperative MRI.

Further software development

We thank BrainLab, Germany for kindly providing the two elastic fusion algorithms used in this study. The software applied in the present study was constantly refined. The results presented here do not necessarily represent the abilities of the most recent software version, as experiences and results driven from our work directly influenced the further development of the software.

298 REFERENCES

- 299 1. Vadera S, Mullin J, Bulacio J, Najm I, Bingaman W, Gonzalez-Martinez J.
300 Stereoelectroencephalography following subdural grid placement for difficult to localize
301 epilepsy. *Neurosurgery*. 2013;72(5):723-9; discussion 729.
302 doi:10.1227/NEU.0b013e318285b4ae.
- 303 2. Gonzalez-Martinez J, Mullin J, Vadera S, et al. Stereotactic placement of depth electrodes
304 in medically intractable epilepsy. *J Neurosurg*. 2014. doi:10.3171/2013.11.JNS13635.
- 305 3. Bancaud J, Talairach J. *La Stéréo-ÉlectroEncéphaloGraphie dans l'épilepsie*. Paris:
306 Masson & Cie; 1965:35-122.
- 307 4. Wyler AR, Ojemann GA, Lettich E, Ward AA. Subdural strip electrodes for localizing
308 epileptogenic foci. *J Neurosurg*. 1984;60(6):1195-200. doi:10.3171/jns.1984.60.6.1195.
- 309 5. Rosenow F, Lüders H. Presurgical evaluation of epilepsy. *Brain*. 2001;124(Pt 9):1683-700.
310 Available at: <http://www.ncbi.nlm.nih.gov/pubmed/11522572>. Accessed August 6, 2013.
- 311 6. Bhavaraju NC, Nagaraddi V, Chetlapalli SR, Osorio I. Electrical and thermal behavior of
312 non-ferrous noble metal electrodes exposed to MRI fields. *Magn Reson Imaging*.
313 2002;20(4):351-7. Available at: <http://www.ncbi.nlm.nih.gov/pubmed/12165354>.
314 Accessed August 12, 2013.
- 315 7. Hastreiter P, Rezk-Salama C, Soza G, et al. Strategies for brain shift evaluation. *Med*
316 *Image Anal*. 2004;8(4):447-64. doi:10.1016/j.media.2004.02.001.
- 317 8. Hermes D, Miller KJ, Noordmans HJ, Vansteensel MJ, Ramsey NF. Automated
318 electrocorticographic electrode localization on individually rendered brain surfaces. *J*
319 *Neurosci Methods*. 2010;185(2):293-8. doi:10.1016/j.jneumeth.2009.10.005.
- 320 9. Serra C, Huppertz H-J, Kockro RA, et al. Rapid and accurate anatomical localization of
321 implanted subdural electrodes in a virtual reality environment. *J Neurol Surg A Cent Eur*
322 *Neurosurg*. 2013;74(3):175-82. doi:10.1055/s-0032-1333124.
- 323 10. Dalal SS, Edwards E, Kirsch HE, Barbaro NM, Knight RT, Nagarajan SS. Localization of
324 neurosurgically implanted electrodes via photograph-MRI-radiograph coregistration. *J*
325 *Neurosci Methods*. 2008;174(1):106-15. doi:10.1016/j.jneumeth.2008.06.028.
- 326 11. Miyagi Y, Shima F, Sasaki T. Brain shift: an error factor during implantation of deep brain
327 stimulation electrodes. *J Neurosurg*. 2007;107(5):989-97. doi:10.3171/JNS-07/11/0989.
- 328 12. King D, Bronen RA, Spencer DD, Spencer SS. Topographic distribution of seizure onset
329 and hippocampal atrophy: relationship between MRI and depth EEG. *Electroencephalogr*
330 *Clin Neurophysiol*. 1997;103(6):692-7. Available at:
331 <http://www.ncbi.nlm.nih.gov/pubmed/9546496>. Accessed January 15, 2014.

- 332 13. Pluim JPW, Maintz JBA, Viergever M a. Mutual-information-based registration of
333 medical images: a survey. *IEEE Trans Med Imaging*. 2003;22(8):986-1004.
334 doi:10.1109/TMI.2003.815867.
- 335 14. Hill DL, Batchelor PG, Holden M, Hawkes DJ. Medical image registration. *Phys Med*
336 *Biol*. 2001;46(3):R1-45. Available at: <http://www.ncbi.nlm.nih.gov/pubmed/11277237>.
337 Accessed August 6, 2013.
- 338 15. Kostelec PJ, Weaver JB, Healy DM. Multiresolution elastic image registration. *Med Phys*.
339 1998;25(9):1593-604. Available at: <http://www.ncbi.nlm.nih.gov/pubmed/9775363>.
340 Accessed August 6, 2013.
- 341 16. Weaver JB, Healy DM, Periaswamy S, Kostelec PJ. Elastic image registration using
342 correlations. *J Digit Imaging*. 1998;11(3 Suppl 1):59-65. Available at:
343 [http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=3453355&tool=pmcentrez&](http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=3453355&tool=pmcentrez&rendertype=abstract)
344 [rendertype=abstract](http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=3453355&tool=pmcentrez&rendertype=abstract). Accessed August 6, 2013.
- 345 17. R Development Core Team. *R: A language and environment for statistical computing*.
346 Vienna, Austria; 2008. Available at: <http://www.r-project.org>.
- 347 18. Wellmer J, von Oertzen J, Schaller C, et al. Digital photography and 3D MRI-based
348 multimodal imaging for individualized planning of resective neocortical epilepsy surgery.
349 *Epilepsia*. 2002;43(12):1543-50. Available at:
350 <http://www.ncbi.nlm.nih.gov/pubmed/12460257>. Accessed August 12, 2013.
- 351 19. Mahvash M, König R, Wellmer J, Urbach H, Meyer B, Schaller K. Coregistration of
352 digital photography of the human cortex and cranial magnetic resonance imaging for
353 visualization of subdural electrodes in epilepsy surgery. *Neurosurgery*. 2007;61(5 Suppl
354 2):340-4; discussion 344-5. doi:10.1227/01.neu.0000303992.87987.17.
- 355 20. Tao JX, Hawes-Ebersole S, Baldwin M, Shah S, Erickson RK, Ebersole JS. The accuracy
356 and reliability of 3D CT/MRI co-registration in planning epilepsy surgery. *Clin*
357 *Neurophysiol*. 2009;120(4):748-53. doi:10.1016/j.clinph.2009.02.002.
- 358 21. LaViolette PS, Rand SD, Ellingson BM, et al. 3D visualization of subdural electrode shift
359 as measured at craniotomy reopening. *Epilepsy Res*. 2011;94(1-2):102-9.
360 doi:10.1016/j.eplesyres.2011.01.011.
- 361 22. Pieters TA, Conner CR, Tandon N. Recursive grid partitioning on a cortical surface model:
362 an optimized technique for the localization of implanted subdural electrodes. *J Neurosurg*.
363 2013;118(5):1086-97. doi:10.3171/2013.2.JNS121450.
- 364 23. Dykstra AR, Chan AM, Quinn BT, et al. Individualized localization and cortical surface-
365 based registration of intracranial electrodes. *Neuroimage*. 2012;59(4):3563-70.
366 doi:10.1016/j.neuroimage.2011.11.046.

FIGURE LEGENDS

Figure 1: Three-dimensional deformation map of elastic image fusion between an MRI and a CT.

Green color shows areas of minor deformation whereas red color indicates those with strong deformation. A) Axial section at the level of the temporal lobes. The artifact of subdurally implanted electrode contacts is shown (b). B) Axial section at the level of the frontal lobes. A relevant right frontal pneumocephalus and its effect on the brain tissue are shown (a).

Figure 2: Distance between electrode contact center and cortex surface

The figure shows a highly effective reduction of the distance between the electrode contact centers and the cortical surface, from a median of 2.3 mm to 0 mm after AEF and GEF in the frontal region. Reductions were also measured in the temporal and parietal regions. All reductions are significant (GEF in temporal region: $p = .03$; all others: $p < .001$). Thick lines indicate medians, boxes indicate the interquartile range, and whiskers indicate the ± 1.58 interquartile range. Outliers are shown as open circles.

Figure 3: Projection of electrode contacts on the brain surface after linear and elastic image fusion

The figure shows the temporal electrode contacts of two illustrative cases after linear (A and C) and after elastic fusion (B and D). After linear fusion some of the contacts are virtually buried in the tissue. Elastic fusion provides a more plausible projection of the contacts onto the brain surface.

linear fusion

A

B

C

D

patient	laterality	region	number of contacts
1	left	temporo-polar	6
		posterior temporal	6
	right	temporo-polar	6
		anterior temporal	6
		posterior temporal	6
2	left	temporo-polar	4
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
	right	temporo-polar	4
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
3	left	temporo-polar	6
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
	right	temporo-polar	6
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
4	left	temporo-polar	6
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
		inferior frontal	6
		superior frontal	8
		fronto-parietal	8
		parietal	8
	right	temporo-polar	6
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
		inferior frontal	6
		superior frontal	8
		fronto-parietal	8
		parietal	8
5	left	temporo-polar	4
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
	right	temporo-polar	4
		anterior temporal	6
		medial temporal	6
		posterior temporal	6
Total number of contacts implanted			274

Table 1: Electrode contacts implanted

The table shows the positions and numbers of contacts per electrode implanted in all five patients.

	No correction	AEF	GEF
Median (mm)	1.90	0	0
IQR	0.75	0.4	0
P		.14 (LF/AEF)	.07 (LF/GEF)
P			.39 (AEF/GEF)

Table 2: Midline shift correction

The table shows the midline shift measured at the septum pellucidum after AEF and GEF and the results compared with the non-corrected CT.

	No correction	AEF
Median	9.8	9.2
IQR	5.2	4.88
p		.68

Table 3: Reduction of frontal pneumencephalus measured as cortex-bone distance

The table shows the maximum thickness of frontal intracranial air measured in an axial slice. The AEF did not lead to a sufficient correction of the brain shift effect.

	No correction	Elastic fusion
Median (mm)	19.49	19.1
IQR	20.7	17.18
p		.86

Table 4: Volumetric reduction of frontal pneumencephalus

As the measurement of the pneumencephalus' thickness the volumetric measurement does not show a satisfying correction of the brain shift effect.

Region	all		frontal		temporal		parietal	
	AEF	GEF	AEF	GEF	AEF	GEF	AEF	GEF
Median	2.9	2.8	2.45	2.45	2.8	2.6	4.3	4.5
IQR	2.53	2.5	2.1	1.88	2.53	2.45	1.8	2.45
p	<div><div></div><div></div><div>.55</div></div>		<div><div></div><div></div><div>.99</div></div>		<div><div></div><div></div><div>.42</div></div>		<div><div></div><div></div><div>.96</div></div>	
			<div><div></div><div></div><div>.031</div></div>					
			<div><div></div><div></div><div><.001</div></div>					
			<div><div></div><div></div><div><.001</div></div>					

Table 5: Relative correction of electrode contact position by elastic fusion

Elastic fusion led to visualization of the electrode contacts in relevant distance to their original position in the postoperative CT. AEF and GEF achieved nearly identical shifts in all regions. The shift was significantly higher in the temporal than in the frontal region and most in the parietal region.