



**University of
Zurich**^{UZH}

**Zurich Open Repository and
Archive**

University of Zurich
University Library
Strickhofstrasse 39
CH-8057 Zurich
www.zora.uzh.ch

Year: 2018

Galvanic coupling of steel and gold alloy lingual brackets with orthodontic wires: Is corrosion a concern?

Polychronis, Georgios ; Al Jabbari, Youssef S ; Eliades, Theodore ; Zinelis, Spiros

DOI: <https://doi.org/10.2319/092917-655.1>

Posted at the Zurich Open Repository and Archive, University of Zurich

ZORA URL: <https://doi.org/10.5167/uzh-169361>

Journal Article

Published Version

Originally published at:

Polychronis, Georgios; Al Jabbari, Youssef S; Eliades, Theodore; Zinelis, Spiros (2018). Galvanic coupling of steel and gold alloy lingual brackets with orthodontic wires: Is corrosion a concern? *Angle Orthodontist*, 88(4):450-457.

DOI: <https://doi.org/10.2319/092917-655.1>

Galvanic coupling of steel and gold alloy lingual brackets with orthodontic wires: *Is corrosion a concern?*

Georgios Polychronis^a; Youssef S. Al Jabbari^b; Theodore Eliades^c; Spiros Zinelis^d

ABSTRACT

Objectives: The aim of this research was to assess galvanic behavior of lingual orthodontic brackets coupled with representative types of orthodontic wires.

Materials and Methods: Three types of lingual brackets: Incognito (INC), In-Ovation L (IOV), and STb (STB) were combined with a stainless steel (SS) and a nickel-titanium (NiTi) orthodontic archwire. All materials were initially investigated by scanning electron microscopy / x-ray energy dispersive spectroscopy (SEM/EDX) while wires were also tested by x-ray diffraction spectroscopy (XRD). All bracket-wire combinations were immersed in acidic 0.1M NaCl 0.1M lactic acid and neutral NaF 0.3% (wt) electrolyte, and the potential differences were continuously recorded for 48 hours.

Results: The SEM/EDX analysis revealed that INC is a single-unit bracket made of a high gold (Au) alloy while IOV and STB are two-piece appliances in which the base and wing are made of SS alloys. The SS wire demonstrated austenite and martensite iron phase, while NiTi wire illustrated an intense austenite crystallographic structure with limited martensite. All bracket wire combinations showed potential differences below the threshold of galvanic corrosion (200 mV) except for INC and STB coupled with NiTi wire in NaF media.

Conclusions: The electrochemical results indicate that all brackets tested demonstrated galvanic compatibility with SS wire, but fluoride treatment should be used cautiously with NiTi wires coupled with Au and SS brackets. (*Angle Orthod.* 2018;88:450–457.)

KEY WORDS: Galvanic corrosion; Brackets; Electrochemistry

INTRODUCTION

Corrosion of orthodontic appliances and the biological consequences are a primary concern of clinical orthodontists, researchers, academics and manufacturers. Therefore, a lot of research has been focused on in vivo phenomena, material characterization, and development of new products involving orthodontic brackets and wires.^{1–6} Primary testing of electrochemical properties of orthodontic alloys used in manufacturing orthodontic devices is of paramount importance. In addition, galvanic coupling between dissimilar alloys, such as brackets and wires, is also a hot topic in orthodontic research as it may trigger galvanic corrosion under clinical conditions, increasing ionic release and degrading the functionality of orthodontic devices.^{7–14}

Lingual orthodontic brackets were introduced in clinical practice in the late 1970s¹⁵ to satisfy the demand for invisible treatment. However, the position-associated patient intolerance,^{16–18} and the in-

^a Research Fellow, Department of Biomaterials, School of Dentistry, National and Kapodistrian University of Athens, Greece.

^b Director, Dental Biomaterials Research and Development Chair, and Professor, Prosthetic Dental Sciences Department, College of Dentistry, King Saud University, Riyadh, Saudi Arabia.

^c Professor and Director, Clinic of Orthodontics and Pediatric Dentistry, Center of Dental Medicine, University of Zurich, Zurich, Switzerland.

^d Associate Professor, Department of Biomaterials, School of Dentistry, National and Kapodistrian University of Athens, Athens, Greece, and Consultant, Dental Biomaterials Research and Development Chair, King Saud University, Riyadh, Saudi Arabia.

Corresponding author: Theodore Eliades, Clinic of Orthodontics and Pediatric Dentistry, Center of Dental Medicine, University of Zurich, Plattenstrasse 11, Zurich 8032, Switzerland (e-mail: theodore.eliades@zzm.uzh.ch)

Accepted: January 2018. Submitted: September 2017.

Published Online: March 6, 2018

© 2018 by The EH Angle Education and Research Foundation, Inc.

creased shape variability of the anterior tooth lingual surfaces¹⁹ partially undermined their usability. Therefore, an effort was made to overcome these two major drawbacks by manufacturing low-profile brackets and customizing them for each patient, respectively. Recently, gold (Au) base brackets have been introduced with big differences in mechanical properties compared with conventional lingual brackets, which are still made of stainless steel (SS) alloys.²⁰ In particular, Au alloy brackets are advertised as highly anticorrosive and the first choice for patients allergic to nickel (Ni). A recent retrieval analysis study pointed out that Au base brackets demonstrated elemental and mechanical stability over the course of orthodontic treatment.²¹ However concerns have arisen related to the big difference in electronegativity between Au alloy and orthodontic wire made of SS and nickel-titanium (NiTi).

The aim of this study was to investigate the galvanic potential of lingual orthodontic brackets combined with representative types of orthodontic wires in different electrolyte media containing chlorides and fluorides.

MATERIALS AND METHODS

Materials Tested

Three types of orthodontic central incisor lingual brackets of 0.022" slot size were included in this study: Incognito (INC) (3M Unitek, Monrovia, Calif, USA), InOvation L (IOV) (Dentsply GAC, Bohemia, NY, USA), and STb (STB) (Light Lingual System, Ormco, Orange, Calif, USA). One rectangular 0.017 × 0.025" stainless steel (SS) orthodontic archwire (SAW) (Steel Arch Wires, Forestadent, Bernhard Forster, Germany), and a Ni-Ti wire of similar dimensions (NSE) (Neo Sentalloy, Dentsply GAC) were also tested.

Scanning Electron Microscopy/X-Ray Energy Dispersive Spectroscopy

Five brackets and five 10-mm segments from the central part of the orthodontic wires were imaged by a scanning electron microscope (QUANTA 200, FEI, Hillsboro, Ore, USA). Secondary electron images were recorded from the surfaces of brackets and wires under the following operational conditions: 10⁻⁶ Torr chamber pressure (high vacuum mode), 20 kV accelerating voltage, 103 μA beam current, and ×100 nominal magnification. The elemental composition of the devices tested was determined using x-ray energy dispersive spectroscopy (EDX) with an X Flash 6|10 Silicon Drift Detector (Bruker, Berlin, Germany). One spectrum was collected from different locations on bracket surfaces (ie, base, soldering alloy, wing) and

orthodontic wire using 20 KV accelerating voltage 107 μA beam current and spot analysis mode.

X-Ray Diffraction Spectroscopy

Ten segments (about 2 mm in length) from each orthodontic wire were retained together along their long axis and embedded in epoxy resin (Epofix, Struers, Belarup, Denmark). Then, the samples were ground up to 4000 grit SiC paper under water cooling and polished with diamond pastes up to 1 μm in a grinding/polishing machine (Dap-V, Struers). Their crystallographic structure was determined using x-ray diffraction spectroscopy (XRD) analysis (D8 Advance; Bruker, Billerica, Mass, USA). Spectra were recorded employing 40 kV accelerating voltage, 40 mA beam current, 30°–100° 2theta angle scan range, 0.02 °/s scanning speed, 0.02° scan step and 1-second preset time.

Electrochemical Testing

Adequate length SS metallic rods were laser welded (R102915, LaserStar Technologies, Riverside, RI, USA) to the bases of the lingual orthodontic brackets. The orthodontic archwires, the bases of the brackets and a portion of the welded rods were insulated with nonconducting epoxy resin and an elastic rubber tube on top. The electrical connection between the two ends (bracket and wire) was verified by a multimeter. The galvanic cell apparatus consisted of a glass container for the electrolyte, a Plexiglas top permitting the entrance of an orthodontic bracket-archwire couple and preventing evaporation of electrolyte, a voltmeter (P903, Consort, Turnhout, Belgium) and a personal computer. The construction of the electrochemical cell and the procedure for measuring galvanic potential difference followed the ASTM G71-81 standard.²² The ratio of exposed surfaces to the electrolytes corresponded as much as possible to the clinical conditions (8 mm length for orthodontic archwires⁷ and the total lingual surface area for the bracket). The volume of electrolyte was selected according to the proposed ratio of 40 mL/cm² metal surface area (≈55 mL). Measurement of potential difference was conducted at room temperature (25°C) three times for each bracket-wire pair using new specimens each time. Data regarding time and potential were collected continuously for 48 hours every 10 seconds since pilot studies showed that the plateau phase was reached much earlier. The galvanic potential at 48 hours was averaged. Before each measurement, the exposed surfaces were cleaned with acetone to prevent organic contamination. The measurement of galvanic potential difference was measured using Consort nv Software

(2002), and the data were graphically represented accounting for the noise as well.

Two electrolytes were used containing chloride and fluoride ions, respectively. The first electrolyte was an acidic (pH = 2.3) water-based solution of lactic acid 0.1M and NaCl 0.1M according to ISO 10271.²³ The second electrolyte was a water-based solution of NaF 0.3% that corresponded to 1394 parts per million (ppm) fluoride ion with a pH = 6.5. Deionized water was used to prepare both electrolytes.

RESULTS

Scanning Electron Microscopy/X-Ray Energy Dispersive Spectroscopy

Figure 1 demonstrates representative SEM images and EDX spectra from the surfaces of the brackets and wires tested. The elemental content of material tested after EDX analysis is presented in Table 1. Figure 1A shows the INC brackets as a single-unit appliance made by a precious alloy (Au-Cu-Pt-Ag) (Figure 1B, Table 1). The IOV consists of a base and wing joined together by an intermediate zone with successive overlapping laser spots (Figure 1C). The EDX analysis (Figure 1D) illustrated that the base and wing were made of different SS alloys. The STB was also a two-piece structure made by a base and wing of different SS alloys joined together by a high Au soldering alloy (Table 1). SAW (Figure 1G) and NSE (Figure 1I) were typical orthodontic wires made of SS and NiTi, respectively (Table 1), with the latter showing a rougher surface compared with the former.

X-Ray Diffraction Spectroscopy

The XRD analysis (Figure 2) showed that SAW consisted of austenite (γ phase) and martensite (α' phase) iron structures. NSE exhibited predominately an austenitic NiTi structure with limited content of martensite.

Electrochemical Testing

The potential differences over time for all bracket-archwire combinations tested are presented in Figure 3. Table 2 demonstrates the mean values and standard deviations of the final potential differences after 48 hours. In all cases, the potential difference seemed to reach a plateau during the first hours after immersion except for STB-NiTi in NaCl/lactic acid and for all bracket-NiTi combinations in NaF, where a constant value was reached at a later stage of recording (Figure 3D). INC coupled with SS wire (SAW) showed the lowest potential difference (-19mV) in the acidic conditions of NaCl/lactic acid solution followed by IOV (-24 mV) and STB (-58 mV), while the potential

differences were eliminated in the NaF media (Table 2). INC and IOV showed minimal differences (4 and -9 mV respectively) when coupled with NiTi wire (NSE) in NaCl/lactic acid solution, while STB showed a higher value (-95 mV). In NaF solution, IOV showed the lowest difference (55 mV) followed by STB (203 mV) and INC (319 mV).

DISCUSSION

Both ferrous brackets (IOV and STB) are two-piece structures that consist of different alloys in the base and wing parts. However, both brackets share similar compositions for base and wing components. The elemental composition for the base fits well with the nominal composition of austenitic SS AISI 316 alloy (%wt: Fe balance, Cr: 16–18, Ni: 10–14, Mo: 2–3, Mn: 2.0),¹⁴ while the wings correspond to the nominal composition of precipitation hardening SS alloy PH 17-4 (%wt: Fe balance, Cr: 15–17, Ni: 3–5, Cu: 3–5, Si: 1, Mn: 1, and traces of P, S, and C).¹³ However, the base and wing are joined by laser welding for IOV, while for STB they are soldered by a high Au soldering alloy. Both technologies (laser welding and soldering) have been extensively used in bracket technology, with the latter always raising concerns for galvanic actions of dissimilar alloys.^{1,13,24} In contrast, INC is a single-piece appliance made of a high Au-Cu-Pt-Ag alloy (Table 1). The XRD analysis of orthodontic wires was essential, especially for Ni-Ti wires as their electrochemical properties are dominated by the presence of different phases.²⁵ The results of XRD analysis in accordance with previously reported XRD results,²⁶ where NSE consisted mainly of an austenitic phase. SAW showed a smoother surface, a finding that is analogous to previously published data.²⁷

Two solutions were used in this study. The first solution was acidic with a pH of 2.3 and is proposed by ISO 10271 for screening testing of electrochemical properties of dental materials.²³ The susceptibility of orthodontic brackets and wires to F ion is concentration-related and has been tested many times in NaF solutions with similar concentrations (0.1~0.5 %wt).^{28–31} Furthermore, the concentration of 0.3% NaF used corresponds to 1394 ppm F ion, which is below the upper limit of 1450 ppm present in child toothpastes.³²

In the presence of acidic NaCl/lactic acid electrolyte, INC brackets exhibited the lowest galvanic values independent of the archwire type (Table 1). Therefore, it seems that the nobility difference between the orthodontic alloys is not the main factor affecting the results, but rather, the passivation dynamics are more important. This involves the formation of chromium^{33,34} and titanium oxide^{35–38} protective layers as well as the tendency to dissolve by the electrolyte constituents




Figure 1. Representative secondary electron images (SEI) and EDX spectra from the surfaces of brackets and wires tested. INC is a single-unit bracket (A), while IOV(C) and STB(E) demonstrate an intermediate zone with successive laser spots and soldering alloy, respectively. EDX spectra show that, for both brackets, bases and wings are made of different SS alloys (D and F). SAW (G) illustrates a smoother surface compared with NSE (I) orthodontic wire. Nominal magnification 100 \times , scale 500 μ m.

Table 1. Mean Values of Elemental Content of All Material Tested After X-Ray Energy Dispersive Spectroscopy Analysis (n = 5)^{a,b}

Elements	INC	IOV		STB			SAW	NSE
		Base	Wing	Base	Soldering	Wing		
Fe		67.5	73.2	71.1	12.2	76.9	72.2	
Cr		17.3	16.3	16.4	3.5	15.5	18.2	
Ni		11.3	3.8	9.9	7.8	4.1	7.6	52.5
Cu	13.2		4		1.7	3.5		
Mo		2.5		2.1			0.4	
Mn		1	0.8	0.6	0.5	0.4	0.8	
Si		0.7	0.7	0.4	0.3	0.5	0.6	
Au	69.6				73.8			
Ag	7.4							
Pt	8.3							
Zn	1.2							
Ti								47.5
Ir	0.6							
Nb			1.0					

^a Standard deviations are less than 0.6 for all elements tested and have been omitted for the sake of clarity.

^b INC indicates Incognito bracket (3M Unitek, Monrovia, Calif); IOV, InOvation L bracket (Dentsply GAC, Bohemia, NY); NSE, nickel-titanium archwire (Neo Sentalloy, Dentsply GAC); SAW, stainless steel archwire (Steel Arch Wires, Forestadent, Bernhard Forster, Germany); STB, STB bracket (Light Lingual System,Ormco, Orange, Calif).

and acidity. In general, SS alloys are more susceptible to chloride ion attack than titanium,³⁰ and acidity acts in a deleterious way synergistically.³⁹ Although the base and wing of IOV and STB are made of the same alloy, they showed different galvanic potentials. The latter showed higher galvanic potential (Table 1) with both wires tested, a finding that should be attributed to the Au-base soldering alloy. Previous studies have reported the development of an intermediate diffusion zone depleted from Au, and thus, it is of questionable corrosion resistance.²⁴ From a clinical point of view, an opposite relationship might be more favorable since brackets remain intraorally throughout treatment, whereas archwires can be replaced monthly or sooner. Nevertheless, the potential difference of all bracket-archwire couples did not surpass 200 mV, which is

considered the threshold for galvanic action between dissimilar alloys. Thus, galvanic corrosion is not anticipated to occur at this particular corrosive environment.

In contrast to NaCl/lactic acid electrolyte, INC brackets were characterized by the worst behavior in combination with NiTi alloy archwires in NaF solution. In a similar manner, the STB bracket, due to its Au brazing, exhibited potential difference close to the limit of 200 mV. Clinically, the ability of the orthodontist to change the NiTi archwires monthly minimizes the alterations to their mechanical properties, but this is not true for the biocompatibility consequences. The concentration of electrolyte fluoride ions used was close to the upper limit of 1450 ppm present in toothpastes for children. Nevertheless, possible fluo-




Figure 2. XRD spectra of SAW (A) and NSE (B) wires tested. (A) SAW consists of austenite (γ phase) and martensite (α' phase) iron structure. (B) NEO shows the presence of NiTi austenite (A) structure with limited presence of martensite (M).




Figure 3. Graphic representation of galvanic potential difference changes recorded for 48 hours in different electrolytic environments. Lingual orthodontic brackets coupled with SS archwire (SAW) in NaCl/lactic acid (A) and NaF (B) media. Lingual brackets coupled with NiTi archwire in NaCl/lactic acid (C) and NaF (D).

ride gel topical therapy, where fluoride concentrations may reach up to 12500 ppm, should be postponed in the presence of NiTi orthodontic archwires, or the wires should be temporarily removed during the process and replaced at the end by the clinician. In any case, low-concentration fluoride dentifrices and toothpastes should be preferred.

A limitation of this study was that the galvanic potentials cannot be extrapolated directly to intraoral conditions because of the multiplicity of factors that could alter the reactivity of materials in the oral environment. These factors include the adsorption of integuments and formation of biofilm, temperature and

pH fluctuations, as well as enzymatic and microbial activity. Contrary to other electrochemical phenomena, such as ionic release in which saliva samples could be obtained and analyzed, experimental testing is the only way to measure the intensity of galvanic coupling between dissimilar alloys. Although the experimental results might differ from what happens under clinical conditions, the ranking of different combinations is indicative of the most vulnerable pairs of materials.

CONCLUSIONS

- The electrochemical results indicate that all brackets tested demonstrated galvanic compatibility with SS

Table 2. Galvanic Coupling Potential Difference Results Recorded After 48 Hours: Mean (SD) of Three Measurements^{a,b}

	Galvanic Potential (mV)			
	SAW		NSE	
	NaCl/Lactic Acid	NaF	NaCl/Lactic Acid	NaF
INC	-19 (16)	0 (2)	4 (2)	319 (44)
IOV	-24 (12)	-1 (1)	-9 (1)	55 (42)
STB	-58 (13)	0 (2)	-95 (23)	203 (2)

^a Positive values correspond to archwire oxidation and bracket reduction while negative to the opposite effect.

^b INC indicates Incognito bracket (3M Unitek, Monrovia, Calif); IOV, InOvation L bracket (Dentsply GAC, Bohemia, NY); NSE, nickel-titanium archwire (Neo Sentalloy, Dentsply GAC); SAW, stainless steel archwire (Steel Arch Wires, Forestadent, Bernhard Forster, Germany); STB, STB bracket (Light Lingual System, Ormco, Orange, Calif).

wire, but caution is suggested in performing fluoride treatment in the case of NiTi wires coupled with Au and SS brackets.

ACKNOWLEDGMENT

The authors extend their appreciation to the International Scientific Partnership Program (ISPP) at King Saud University for funding this research work through ISPP 0060.

REFERENCES

- Eliades T, Zinelis S, Bouraueul C, Eliades G. Manufacturing of orthodontic brackets: a review of metallurgical perspectives and applications. *Recent Pat Mater Sci*. 2008;1:135–139.
- Mikulewicz M, Chojnacka K. Release of metal ions from orthodontic appliances by in vitro studies: a systematic literature review. *Biol Trace Elem Res*. 2011;139:241–256.
- Mikulewicz M, Chojnacka K, Wozniak B, Downarowicz P. Release of metal ions from orthodontic appliances: an in vitro study. *Biol Trace Elem Res*. 2012;146:272–280.
- Mikulewicz M, Chojnacka K, Zielinska A, Michalak I. Exposure to metals from orthodontic appliances by hair mineral analysis. *Environ Toxicol Pharmacol*. 2011;32:10–16.
- Eliades T, Bouraueul C. Intraoral aging of orthodontic materials: the picture we miss and its clinical relevance. *Am J Orthod Dentofacial Orthop*. 2005;127:403–412.
- House K, Sernetz F, Dymock D, Sandy JR, Ireland AJ. Corrosion of orthodontic appliances—should we care? *Am J Orthod Dentofacial Orthop*. 2008;133:584–592.
- Bakhtari A, Bradley TG, Lobb WK, Berzins DW. Galvanic corrosion between various combinations of orthodontic brackets and archwires. *Am J Orthod Dentofacial Orthop*. 2011;140:25–31.
- Tahmasbi S, Ghorbani M, Masudrad M. Galvanic corrosion of and ion release from various orthodontic brackets and wires in a fluoride-containing mouthwash. *J Dent Res Dent Clin Dent Prospects*. 2015;9:159–165.
- Iijima M, Endo K, Yuasa T, et al. Galvanic corrosion behavior of orthodontic archwire alloys coupled to bracket alloys. *Angle Orthod*. 2006;76:705–711.
- Varma DP, Chidambaram S, Reddy KB, Vijay M, Ravindranath D, Prasad MR. Comparison of galvanic corrosion potential of metal injection molded brackets to that of conventional metal brackets with nickel-titanium and copper nickel-titanium archwire combinations. *J Contemp Dent Pract*. 2013;14:488–495.
- Nayak RS, Shafiuddin B, Pasha A, Vinay K, Narayan A, Shetty SV. Comparison of galvanic currents generated between different combinations of orthodontic brackets and archwires using potentiostat: an in vitro study. *J Int Oral Health*. 2015;7:29–35.
- Schiff N, Boinet M, Morgon L, Lissac M, Dalard F, Grosogeat B. Galvanic corrosion between orthodontic wires and brackets in fluoride mouthwashes. *Eur J Orthod*. 2006;28:298–304.
- Siargos B, Bradley TG, Darabara M, Papadimitriou G, Zinelis S. Galvanic corrosion of metal injection molded (MIM) and conventional brackets with nickel-titanium and copper-nickel-titanium archwires. *Angle Orthod*. 2007;77:355–360.
- Darabara MS, Bourithis LI, Zinelis S, Papadimitriou GD. Metallurgical characterization, galvanic corrosion, and ionic release of orthodontic brackets coupled with Ni-Ti archwires. *J Biomed Mater Res B Appl Biomater*. 2007;81:126–134.
- Fujita K. New orthodontic treatment with lingual bracket mushroom arch wire appliance. *Am J Orthod*. 1979;76:657–675.
- Long H, Zhou Y, Pyakurel U, et al. Comparison of adverse effects between lingual and labial orthodontic treatment. *Angle Orthod*. 2013;83:1066–1073.
- Wu A, McGrath C, Wong RW, Wiechmann D, Rabie AB. Comparison of oral impacts experienced by patients treated with labial or customized lingual fixed orthodontic appliances. *Am J Orthod Dentofacial Orthop*. 2011;139:784–790.
- Caniklioglu C, Ozturk Y. Patient discomfort: a comparison between lingual and labial fixed appliances. *Angle Orthod*. 2005;75:86–91.
- McIntyre GT, Millett DT. Crown-root shape of the permanent maxillary central incisor. *Angle Orthod*. 2003;73:710–715.
- Zinelis S, Sifakakis I, Katsaros C, Eliades T. Microstructural and mechanical characterization of contemporary lingual orthodontic brackets. *Eur J Orthod*. 2014;36:389–393.
- Hersche S, Sifakakis I, Zinelis S, Eliades T. Elemental, microstructural, and mechanical characterization of high gold orthodontic brackets after intraoral aging. *Biomed Tech (Berl)* 2017;62:97–102.
- ASTM:G71-81. *Standard Guide for Conducting and Evaluating Galvanic Corrosion Tests in Electrolytes*. West Conshohocken, PA: ASTM International; 2014.
- ISO:10271. *Dentistry-Corrosion Test Methods for Metallic Materials*. Geneva, Switzerland: International Organization for Standardization; 2011.
- Zinelis S, Annousaki O, Eliades T, Makou M. Elemental composition of brazing alloys in metallic orthodontic brackets. *Angle Orthod*. 2004;74:394–399.
- Pun DK, Berzins DW. Corrosion behavior of shape memory, superelastic, and nonsuperelastic nickel-titanium-based orthodontic wires at various temperatures. *Dent Mater*. 2008;24:221–227.
- Brantley W. Orthodontic wires. In: Brantley W, Eliades T, eds. *Orthodontic Materials*. Stuttgart, Germany: Thieme; 2001:78–100.
- Huang HH. Corrosion resistance of stressed NiTi and stainless steel orthodontic wires in acid artificial saliva. *J Biomed Mater Res A*. 2003;66:829–839.
- Kao CT, Ding SJ, He H, Chou MY, Huang TH. Cytotoxicity of orthodontic wire corroded in fluoride solution in vitro. *Angle Orthod*. 2007;77:349–354.
- Lee TH, Huang TK, Lin SY, Chen LK, Chou MY, Huang HH. Corrosion resistance of different nickel-titanium archwires in acidic fluoride-containing artificial saliva. *Angle Orthod*. 2010;80:547–553.
- Jang HS, Son WS, Park SB, Kim HI, Yong HK. Effect of acetic NaF solution on the corrosion behavior of stainless steel orthodontic brackets. *Dent Mater J*. 2006;25:339–344.
- Li X, Wang J, Han EH, Ke W. Influence of fluoride and chloride on corrosion behavior of NiTi orthodontic wires. *Acta Biomater*. 2007;3:807–815.
- Spencer AJ. The use of fluorides in Australia: guidelines. *Aust Dent J*. 2006;51:195–199.
- Lin MC, Lin SC, Lee TH, Huang HH. Surface analysis and corrosion resistance of different stainless steel orthodontic brackets in artificial saliva. *Angle Orthod*. 2006;76:322–329.

34. Eliades T, Athanasiou AE. In vivo aging of orthodontic alloys: implications for corrosion potential, nickel release, and biocompatibility. *Angle Orthod.* 2002;72:222–237.
35. Kao CT, Huang TH. Variations in surface characteristics and corrosion behaviour of metal brackets and wires in different electrolyte solutions. *Eur J Orthod.* 2010;32:555–560.
36. Daems J, Celis JP, Willems G. Morphological characterization of as-received and in vivo orthodontic stainless steel archwires. *Eur J Orthod.* 2009;31:260–265.
37. Iijima M, Endo K, Ohno H, Yonekura Y, Mizoguchi I. Corrosion behavior and surface structure of orthodontic Ni-Ti alloy wires. *Dent Mater J.* 2001;20:103–113.
38. Rondelli G, Vicentini B. Localized corrosion behaviour in simulated human body fluids of commercial Ni-Ti orthodontic wires. *Biomaterials.* 1999;20:785–792.
39. Huang HH. Surface characterizations and corrosion resistance of nickel-titanium orthodontic archwires in artificial saliva of various degrees of acidity. *J Biomed Mater Res A.* 2005;74:629–639.